



Paragon Analytics

Radiochemistry Case Narrative

Isotopic Polonium

New Horizons

CSMRI / 2135

PA WO 0404241

1. This report consists of the analytical results and supporting documentation for six soil samples received by Paragon on 04/23/04.
2. These samples were prepared according to Paragon Analytics procedures PA SOP773R8 and PA SOP711R1.
3. The samples were analyzed for the presence of Polonium-210 according to Paragon Analytics procedure PA SOP714R8. The analyses were completed on 05/19/04.
4. The isotopic analysis results for these samples are reported on a dry weight basis in units of pCi/gram.
5. Due to the poor availability of a 'NIST traceable' Polonium-209 tracer solution, the concentration of a dilution of isotopically pure Po-209 received from Oak Ridge National Laboratory was determined by cross calibration against two distinct NIST traceable sources (Po-210 and Pb/Po-210). While the calibration is not formally 'NIST traceable', the Po-209 concentration was calibrated against and is thus referenced to NIST traceable standards.
6. Results of this analysis are decay-corrected to the sampling date, based on the 138.4 day half-life of Po-210. This decay correction makes no assumptions as to the equilibrium state of Po-210 with the Pb-210 parent nuclide, which has a half-life of 22.3 years.
7. Paragon Analytics follows the convention outlined in ANSI N42.23 for reporting significant digits in the TPU and MDC results. ANSI N42.23 states that the TPU result should be rounded to two significant digits and that the MDC result should be rounded to the same decimal place as the TPU result. In practice, this could result in an MDC result with a reported value of 0 for samples with significant activity, including the batch laboratory control sample.
8. The requested MDC of 0.1 pCi/gram was not achieved for samples B117 and H07S (PA ID's 0404241-3 and -11). The reportable Po-210 activity in these samples exceeds their achieved MDC's. These samples are identified with an 'M3' flag on the final reports.
9. No further anomalous situations were encountered during the preparation or analysis of these samples. All remaining quality control criteria were met.

000001


The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, Paragon Analytics certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.



John Petrovic
Radiochemistry Instrumentation

5/21/04

Date



Radiochemistry Final Data Review

5/21/04

Date

000002

PARAGON ANALYTICS
Radiochemistry Data Package

Section 1

**SAMPLE RESULTS
SUMMARY**

Polonium-210 by Alpha Spectroscopy Sample Results Summary

Client Name: New Horizons
 Client Project Name: CSMRI
 Client Project Number: 2135

Laboratory Name: Paragon Analytics
 PAI Work Order: 0404241

Page: 1 of 1
 Reported on: Friday, May 21, 2004
 9:41:39 AM

Lab Sample ID	Client Sample ID	Sample Type	Nuclide	Result +/- 2 s TPU	MDC	Units	Matrix	Prep Batch	Date Analyzed	Flags
0404241-3	BI17	Sample	Po-210	6.0 +/- 1.2	0.1	pCi/g	SOIL	PL040513-1	5/19/2004	M3
0404241-5	BI37	Sample	Po-210	15.6 +/- 2.7	0.1	pCi/g	SOIL	PL040513-1	5/19/2004	
0404241-7	BI42	Sample	Po-210	0.56 +/- 0.25	0.06	pCi/g	SOIL	PL040513-1	5/19/2004	
0404241-9	H07	Sample	Po-210	98 +/- 15	0	pCi/g	SOIL	PL040513-1	5/19/2004	
0404241-11	H07S	Sample	Po-210	24.3 +/- 4.1	0.1	pCi/g	SOIL	PL040513-1	5/19/2004	M3
0404241-12	BB04	Sample	Po-210	32.2 +/- 5.4	0.1	pCi/g	SOIL	PL040513-1	5/19/2004	

Comments:

Data Package ID: Po0404241-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- M - The requested MDC was not met.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Date Printed: Friday, May 21, 2004

Paragon Analytics
 LIMS Version: 5.018A

2

PARAGON ANALYTICS
Radiochemistry Data Package

Section 2

**QC RESULTS
SUMMARY**

000005

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Method Blank Results

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Lab ID: PL040513-1MB

Sample Matrix: SOIL

Prep Batch: PL040513-1

Final Aliquot: 0.250 g

Prep SOP: PAI 711 Rev 5

QC Batch ID: PL040513-1-1

Result Units: pCi/g

Date Collected: 13-May-04

Run ID: PL040513-1A

File Name: O40511B

Date Prepared: 13-May-04

Count Time: 300 minutes

Date Analyzed: 19-May-04

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
13981-52-7	Po-210	0.026 +/- 0.093	0.070	U

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	16.81	12.8	pCi/g	76.4	20 - 110 %	

Comments:

Qualifiers/Flags:

U - Result is less than the sample specific MDC.
Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
Y2 - Chemical Yield outside default limits.
LT - Result is less than Requested MDC, greater than sample specific MDC.

Abbreviations:

TPU - Total Propagated Uncertainty (see PAI SOP 743)
MDC - Minimum Detectable Concentration (see PAI SOP 709)
BDL - Below Detection Limit

M - Requested MDC not met.
B - Analyte concentration greater than MDC.
B3 - Analyte concentration greater than MDC but less than Requested MDC.

Data Package ID: PO0404241-1

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Laboratory Control Sample(s)

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Lab ID: PL040513-1LCS	Sample Matrix: SOIL Prep SOP: PAI 711 Rev 5 Date Collected: 13-May-04 Date Prepared: 13-May-04 Date Analyzed: 19-May-04	Prep Batch: PL040513-1 QCBatchID: PL040513-1-1 Run ID: PL040513-1A Count Time: 300 minutes	Final Aliquot: 0.250 g Result Units: pCi/g File Name: O40511L
------------------------------	--	---	--

CASNO	Target Nuclide	Results +/- 2s TPU	MDC	Spike Added	% Rec	Control Limits	Lab Qualifier
13981-52-7	Po-210	32.9 +/- 5.6	0.2	28.4	116	83 - 117	P,M3

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	16.81	11.4	pCi/g	67.6	20 - 110 %	

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- L - LCS Recovery below lower control limit.
- H - LCS Recovery above upper control limit.
- P - LCS Recovery within control limits.
- M - The requested MDC was not met.
- M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)

Data Package ID: PO0404241-1

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Duplicate Sample Results (DER)

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Field ID: BI42
Lab ID: 0404241-7DUP

Sample Matrix: SOIL
Prep SOP: PAI 711 Rev 5
Date Collected: 19-Apr-04
Date Prepared: 13-May-04
Date Analyzed: 19-May-04

Prep Batch: PL040513-1
QCBatchID: PL040513-1-1
Run ID: PL040513-1A
Count Time: 300 minutes
Report Basis: Dry Weight

Final Allquot: 0.260 g
Prep Basis: Dry Weight
Moisture(%): NA
Result Units: pCi/g
File Name: O42417D

CASNO	Analyte	Sample Result +/- 2 s TPU	Duplicate Result +/- 2 s TPU	DER	Control Limit	Lab Qualifiers
13981-52-7	Po-210	0.56 +/- 0.25	0.84 +/- 0.34	0.66	2.13	M3

Comments:

Duplicate Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- W - DER is greater than Warning Limit of 1.42
- D - DER is greater than Control Limit of 2.13
- LT - Result is less than Request MDC, greater than sample specific MDC
- M - Requested MDC not met.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- L - LCS Recovery below lower control limit.
- H - LCS Recovery above upper control limit.
- P - LCS, Matrix Spike Recovery within control limits.
- N - Matrix Spike Recovery outside control limits

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- DER - Duplicate Error Ratio
- BDL - Below Detection Limit
- NR - Not Reported

Data Package ID: PO0404241-1

PARAGON ANALYTICS
Radiochemistry Data Package

3

Section 3

**INDIVIDUAL
SAMPLE RESULTS**

000003

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Sample Results

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Field ID: BI17
Lab ID: 0404241-3

Sample Matrix: SOIL
Prep SOP: PAI 711 Rev 5
Date Collected: 19-Apr-04
Date Prepared: 13-May-04
Date Analyzed: 19-May-04

Prep Batch: PL040513-1
QCBatchID: PL040513-1-1
Run ID: PL040513-1A
Count Time: 300 minutes
Report Basis: Dry Weight

Final Aliquot: 0.267 g
Prep Basis: Dry Weight
Moisture(%): NA
Result Units: pCi/g
File Name: O42413

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
13981-52-7	Po-210	6.0 +/- 1.2	0.1	M3

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	15.74	12.5	pCi/g	79.4	20 - 110 %	

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Data Package ID: PO0404241-1

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Sample Results

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Field ID: B137	Sample Matrix: SOIL	Prep Batch: PL040513-1	Final Aliquot: 0.255 g
Lab ID: 0404241-5	Prep SOP: PAI 711 Rev 5	QC Batch ID: PL040513-1-1	Prep Basis: Dry Weight
	Date Collected: 19-Apr-04	Run ID: PL040513-1A	Moisture(%): NA
	Date Prepared: 13-May-04	Count Time: 300 minutes	Result Units: pCi/g
	Date Analyzed: 19-May-04	Report Basis: Dry Weight	File Name: O42415

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
13981-52-7	Po-210	15.6 +/- 2.7	0.1	

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	16.47	13.2	pCi/g	80.4	20 - 110 %	

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Data Package ID: PO0404241-1

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Sample Results

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Field ID: BI42
Lab ID: 0404241-7

Sample Matrix: SOIL
Prep SOP: PAI 711 Rev 5
Date Collected: 19-Apr-04
Date Prepared: 13-May-04
Date Analyzed: 19-May-04

Prep Batch: PL040513-1
QCBatchID: PL040513-1-1
Run ID: PL040513-1A
Count Time: 300 minutes
Report Basis: Dry Weight

Final Aliquot: 0.255 g
Prep Basis: Dry Weight
Moisture(%): NA
Result Units: pCi/g
File Name: O42417

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
13981-52-7	Po-210	0.56 +/- 0.25	0.06	

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	16.48	15.6	pCi/g	94.6	20 - 110 %	

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Data Package ID: PO0404241-1

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Sample Duplicate Results

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Field ID: BI42
Lab ID: 0404241-7DUP

Sample Matrix: SOIL
Prep SOP: PAI 711 Rev 5
Date Collected: 19-Apr-04
Date Prepared: 13-May-04
Date Analyzed: 19-May-04

Prep Batch: PL040513-1
QCBatchID: PL040513-1-1
Run ID: PL040513-1A
Count Time: 300 minutes
Report Basis: Dry Weight

Final Aliquot: 0.260 g
Prep Basis: Dry Weight
Moisture(%): NA
Result Units: pCi/g
File Name: O42417D

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
13981-52-7	Po-210	0.84 +/- 0.34	0.18	M3

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	16.19	11.8	pCi/g	73.2	20 - 110 %	

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M - The requested MDC was not met.
- M3 - The requested MDC was not met, but thereported activity is greater than the reported MDC.
- W - DER is greater than Warning Limit of 1.42
- D - DER is greater than Control Limit of 2.13

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Data Package ID: PO0404241-1

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Sample Results

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Field ID: H07	Sample Matrix: SOIL	Prep Batch: PL040513-1	Final Allquot: 0.253 g
Lab ID: 0404241-9	Prep SOP: PAI 711 Rev 5	QCBatchID: PL040513-1-1	Prep Basis: Dry Weight
	Date Collected: 22-Apr-04	Run ID: PL040513-1A	Moisture(%): NA
	Date Prepared: 13-May-04	Count Time: 300 minutes	Result Units: pCi/g
	Date Analyzed: 19-May-04	Report Basis: Dry Weight	File Name: O42419

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
13981-52-7	Po-210	98 +/- 15	0	

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	16.62	14.2	pCi/g	85.6	20 - 110 %	

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Data Package ID: PO0404241-1

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Sample Results

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Field ID: H07S
Lab ID: 0404241-11

Sample Matrix: SOIL
Prep SOP: PAI 711 Rev 5
Date Collected: 22-Apr-04
Date Prepared: 13-May-04
Date Analyzed: 19-May-04

Prep Batch: PL040513-1
QC Batch ID: PL040513-1-1
Run ID: PL040513-1A
Count Time: 300 minutes
Report Basis: Dry Weight

Final Aliquot: 0.251 g
Prep Basis: Dry Weight
Moisture(%): NA
Result Units: pCi/g
File Name: O424111

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
13981-52-7	Po-210	24.3 +/- 4.1	0.1	M3

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	16.73	13.9	pCi/g	83.0	20 - 110 %	

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Data Package ID: PO0404241-1

Polonium-210 by Alpha Spectroscopy

PAI 714 Rev 8

Sample Results

Lab Name: Paragon Analytics

Work Order Number: 0404241

Client Name: New Horizons

ClientProject ID: CSMRI 2135

Field ID: BB04	Sample Matrix: SOIL	Prep Batch: PL040513-1	Final Allquot: 0.251 g
Lab ID: 0404241-12	Prep SOP: PAI 711 Rev 5	QC Batch ID: PL040513-1-1	Prep Basis: Dry Weight
	Date Collected: 16-Apr-04	Run ID: PL040513-1A	Moisture(%): NA
	Date Prepared: 13-May-04	Count Time: 300 minutes	Result Units: pCi/g
	Date Analyzed: 19-May-04	Report Basis: Dry Weight	File Name: O424112

CASNO	Target Nuclide	Result +/- 2 s TPU	MDC	Lab Qualifier
13981-52-7	Po-210	32.2 +/- 5.4	0.1	

Chemical Yield Summary

Carrier/Tracer	Amount Added	Result	Units	Yield	Control Limits	Flag
Po-209	16.75	13.8	pCi/g	82.5	20 - 110 %	

Comments:

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative Yield is assumed.
- Y2 - Chemical Yield outside default limits.
- LT - Result is less than Requested MDC, greater than sample specific MDC.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- M - The requested MDC was not met.

Abbreviations:

- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- BDL - Below Detection Limit

Data Package ID: PO0404241-1

PARAGON ANALYTICS
Radiochemistry Data Package

Section 4

4

RAW DATA

000017

Polonium-210 by Alpha Spectroscopy Raw Data Report

Laboratory Name: Paragon Analytics
 PAI Work Order: 0404241

Prep SOP: PAI 711
 Analytical SOP: PAI 714

Reported on: Friday, May 21, 2004
 9:39:58 AM

Sample ID QC Type	Nuclide Type	Sample Date/Time	Prep Batch QC/BatchID	Matrix %Moist.	Decay Date/Time	Ingrrowth Date /Time	Samp Alq Analy Alq	Inst ID Det ID	AnRunID File Name	Count Date/Time	Net Cnts Bkg Cnts	BaseEff Bkg(min)	CntDur(min) Yield	Activity +/- 2 s TPU	MDC DeclEv	ReportUnits ReportBasis	DER RPD	%Spk. Recov Flags
0404241-3 SMP	Po-209 Tracer	4/19/2004 11:00:00 AM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.534 g 0.267 g	2	PL040513-1A O42413	5/19/2004 8:00 AM	696,700 1,000	31.36% 1000	300 79.4%	12.5 2.0	0.1 NA	pCi/g Dry Weight	NA NA	NA
0404241-3 SMP	Po-210 Trg. Analyte	4/19/2004 11:00:00 AM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.534 g 0.267 g	2	PL040513-1A O42413	5/19/2004 8:00 AM	226,700 1,000	31.36% 1000	300 79.4%	6.0 1.2	0.1 NA	pCi/g Dry Weight	NA NA	M3
0404241-5 SMP	Po-209 Tracer	4/19/2004 2:15:00 PM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.511 g 0.255 g	3	PL040513-1A O42415	5/19/2004 8:01 AM	721,400 2,000	32.05% 1000	300 80.4%	13.2 2.1	0.1 NA	pCi/g Dry Weight	NA NA	NA
0404241-5 SMP	Po-210 Trg. Analyte	4/19/2004 2:15:00 PM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.511 g 0.255 g	3	PL040513-1A O42415	5/19/2004 8:01 AM	589,000 0,000	32.05% 1000	300 80.4%	15.6 2.7	0.1 NA	pCi/g Dry Weight	NA NA	NA
0404241-7 SMP	Po-209 Tracer	4/19/2004 3:15:00 PM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.51 g 0.255 g	4	PL040513-1A O42417	5/19/2004 8:01 AM	813,400 2,000	30.72% 1000	300 94.6%	15.6 2.4	0.1 NA	pCi/g Dry Weight	NA NA	NA
0404241-7 SMP	Po-210 Trg. Analyte	4/19/2004 3:15:00 PM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.51 g 0.255 g	4	PL040513-1A O42417	5/19/2004 8:01 AM	24,000 0,000	30.72% 1000	300 94.6%	0.56 0.25	0.06 NA	pCi/g Dry Weight	NA NA	NA
0404241-7 DUP	Po-209 Tracer	4/19/2004 3:15:00 PM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.519 g 0.26 g	5	PL040513-1A O42417D	5/19/2004 8:01 AM	655,200 6,000	31.99% 1000	300 73.2%	11.8 1.9	0.2 NA	pCi/g Dry Weight	NA NA	NA
0404241-7 DUP	Po-210 Trg. Analyte	4/19/2004 3:15:00 PM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.519 g 0.26 g	5	PL040513-1A O42417D	5/19/2004 8:01 AM	29,400 2,000	31.99% 1000	300 73.2%	0.84 0.25	0.18 NA	pCi/g Dry Weight	0.66 NA	M3
0404241-9 SMP	Po-209 Tracer	4/22/2004 1:00:00 PM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.506 g 0.253 g	7	PL040513-1A O42419	5/19/2004 8:02 AM	762,500 5,000	31.80% 1000	300 85.6%	14.2 2.2	0.2 NA	pCi/g Dry Weight	NA NA	NA
0404241-9 SMP	Po-210 Trg. Analyte	4/22/2004 1:00:00 PM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.506 g 0.253 g	7	PL040513-1A O42419	5/19/2004 8:02 AM	3919,000 0,000	31.80% 1000	300 85.6%	98 15	0 NA	pCi/g Dry Weight	NA NA	NA
0404241-11 SMP	Po-209 Tracer	4/22/2004 1:05:00 PM	PL040513-1-1	SOIL NA	5/15/2004 10:00:00 AM	NA	0.502 g 0.251 g	8	PL040513-1A O424111	5/19/2004 8:02 AM	734,400 2,000	31.58% 1000	300 83.0%	13.9 2.2	0.1 NA	pCi/g Dry Weight	NA NA	NA

Comments:

Data Package ID: Po0404241-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- W - DER is greater than Warning Limit of 1.42
- D - DER is greater than Control Limit of 2.13
- + - Duplicate RPD not within limits.
- LT - Result is less than Request MDC, greater than sample specific MDC
- * - Aliquot Basis is 'As Received' while the Report Basis is 'Dry Weight'
- # - Aliquot Basis is 'Dry Weight' while the Report Basis is 'As Received'

Notes:

- 1) The Tracer results are not yield corrected (i.e. activity measured not activity added).
- 2) Where sample time is not available, 12:00 PM (Mountain) is used for decay correction.

Abbreviations:

- TR - Tracer
- TA - Target Analyte
- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- DER - Duplicate Error Ratio
- BDL - Below Detection Limit

Date Printed: Friday, May 21, 2004

Paragon Analytics

LIMS Version: 5.018A

000018

Polonium-210 by Alpha Spectroscopy Raw Data Report

Laboratory Name: Paragon Analytics
PAI Work Order: 0404241

Prep SOP: PAI 711
Analytical SOP: PAI 714

Reported on: Friday, May 21, 2004
9:39:58 AM

Sample ID QC Type	Nuclide Type	Sample Date/Time	Prep Batch QC Batch ID	Ingrowth Date/Time	Decay Date/Time	Matrix %Moist.	Samp. Alq Analy Alq	Inst ID Det ID	AnRunID File Name	Count Date/Time	Net Cnts Bkg Cnts	BaseEff Bkg(min)	Yield	CntDur(min)	Activity +/- 2 s TPU	MDC DeclEv	ReportUnits ReportBasis	DER RPD	%Spk. Recov Flags
0404241-11	Po-210 Trg. Analyte	4/22/2004 1:05:00 PM	PL040513-1	NA	5/15/2004 10:00:00 AM	SOIL	0.502 g 0.251 g	Alpha Spec 8	PL040513-1A O424111	5/19/2004 8:02 AM	931,700 1,000	31.59% 1000	300 83.0%	300	24.3 4.1	0.1 NA	pCi/g Dry Weight	NA NA	M3
0404241-12	Po-209 Tracer	4/16/2004 1:10:00 PM	PL040513-1	NA	5/15/2004 10:00:00 AM	SOIL	0.502 g 0.251 g	Alpha Spec 49	PL040513-1A O424112	5/19/2004 8:03 AM	662,100 3,000	28.66% 1000	300 82.5%	300	13.8 2.2	0.1 NA	pCi/g Dry Weight	NA NA	
0404241-12	Po-210 Trg. Analyte	4/16/2004 1:10:00 PM	PL040513-1	NA	5/15/2004 10:00:00 AM	SOIL	0.502 g 0.251 g	Alpha Spec 49	PL040513-1A O424112	5/19/2004 8:03 AM	1060,000 0,000	28.66% 1000	300 82.5%	300	32.2 5.4	0.1 NA	pCi/g Dry Weight	NA NA	
PL040513-1	Po-209 Tracer	5/13/2004 12:46:37 PM	PL040513-1	NA	5/15/2004 10:00:00 AM	SOIL	0.5 g 0.25 g	Alpha Spec 50	PL040513-1A O40511B	5/19/2004 8:03 AM	671,800 4,000	31.41% 1000	300 76.4%	300	12.8 2.0	0.1 NA	pCi/g Dry Weight	NA NA	
PL040513-1	Po-210 Trg. Analyte	5/13/2004 12:46:37 PM	PL040513-1	NA	5/15/2004 10:00:00 AM	SOIL	0.5 g 0.25 g	Alpha Spec 50	PL040513-1A O40511B	5/19/2004 8:03 AM	1,000 0,000	31.41% 1000	300 76.4%	300	0.026 0.093	0.070 NA	pCi/g Dry Weight	NA NA	U
PL040513-1	Po-209 Tracer	5/13/2004 12:46:37 PM	PL040513-1	NA	5/15/2004 10:00:00 AM	SOIL	0.5 g 0.25 g	Alpha Spec 51	PL040513-1A O40511L	5/19/2004 8:03 AM	585,400 2,000	30.94% 1000	300 67.6%	300	11.4 1.8	0.1 NA	pCi/g Dry Weight	NA NA	
PL040513-1	Po-210 Trg. Analyte	5/13/2004 12:46:37 PM	PL040513-1	NA	5/15/2004 10:00:00 AM	SOIL	0.5 g 0.25 g	Alpha Spec 51	PL040513-1A O40511L	5/19/2004 8:03 AM	1112,700 1,000	30.94% 1000	300 67.6%	300	32.9 5.6	0.2 NA	pCi/g Dry Weight	NA NA	P, M3

Comments:

Data Package ID: Po0404241-1

Qualifiers/Flags:

- U - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- W - DER is greater than Warning Limit of 1.42
- D - DER is greater than Control Limit of 2.13
- + - Duplicate RPD not within limits.
- LT - Result is less than Request MDC, greater than sample specific MDC
- * - Aliquot Basis is 'As Received' while the Report Basis is 'Dry Weight'.
- # - Aliquot Basis is 'Dry Weight' while the Report Basis is 'As Received'.
- M - Requested MDC not met.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- L - LCS Recovery below lower control limit.
- H - LCS Recovery above upper control limit.
- P - LCS, Matrix Spike Recovery within control limits.
- N - Matrix Spike Recovery outside control limits
- NC - Not Calculated for duplicate results less than 5 times MDC
- B - Analyte concentration greater than MDC.
- B3 - Analyte concentration greater than MDC but less than Requested MDC.

Notes:

- 1) The Tracer results are not yield corrected (i.e. activity measured not activity added).
- 2) Where sample time is not available, 12:00 PM (Mountain) is used for decay correction.

Abbreviations:

- TR - Tracer
- TA - Target Analyte
- TPU - Total Propagated Uncertainty (see PAI SOP 743)
- MDC - Minimum Detectable Concentration (see PAI SOP 709)
- DER - Duplicate Error Ratio
- BDL - Below Detection Limit

Date Printed: Friday, May 21, 2004

Paragon Analytics


LIMS Version: 5.018A

Page 2 of 2

00019

042413

AlphaVision Relative Region-Of-Interest (Slope Recalibration)




Acquired: 08:00:46 on 19-May-2004
File: C:\USER\ALPHA\ALPHA\O42413.SPC
Sample: 0404241-3 PL040513-1

Real Time: 18000.48 s. Live Time: 18000.00 s.
Detector: #2 MCB 1 Input 2
Type: PO-210

120000

042415

AlphaVision Relative Region-Of-Interest (Slope Recalibration)



Acquired: 08:01:04 on 19-May-2004

File: C:\User\Alpha\ALPHA\O42415.SPC

Sample: 0404241-5 PL040513-1

Real Time: 18000.48 s. Live Time: 18000.00 s.

Detector: #3 MCB 1 Input 3

Type: PO-210

000023

Paragon Analytics

Alpha Spectroscopy Analysis

Report Printed:
5/19/04 4:45:55 PM

Para0327.rpt
rev 11/13/03 KVG

Sample Name: 0404241-7 PL040513-1

Analysis Type: PO-210

Detector: MCB 1 Input 4

Date/Time of Count: 5/19/04 8:01:23 AM

Sample Volume: 0.510 Total, 0.255 Aliquot.

Live Time: 300.00 Minutes

Chem. Yield: 94.57%

Real Time: 300.01 Minutes

Total Eff.: 29.05 %

Dead Time: 0.0 %

Tracer Amount: 18.666 DPM.

Acquisition: 512 Channels

Efficiency: 30.72%

Analysis: Relative Region-Of-Interest

Original: 2,944 + 10.6222 * Chn + -0.00170 * Chn **2.

Spectrum Calibration: 2,944 + 10.7279 * Chn + -0.00170 * Chn **2.

Cal File:

Spectrum File: C:\User\Alpha\ALPHA\O42417.SPC

Background File: C:\USER\ALPHA\BKGND\B4051804.SPC

Library File: C:\User\Alpha\ALPHAVIS.ALB

Peaks

Peak	Channel	Start	End	FWHM	Height	Gross Cts	Bkg Cts	Net Area	DPM
1	228.32	208	233	10.00	5.00	24.00	0.00	24.00	0.28
Tracer	186.18	169	193	6.00	112.00	814.00	0.60	813.40	9.33

Analysis Results

Peak	Nuclide	Energy (keV)	Width (keV)	Aliquot pCi	MDA pCi	% Error
1	Po-210	5304.38	99.52	0.486	n/a	40.01 %
Tracer	Po-209	4882.00	60.57	16.486	n/a	6.87 %

Totals

		% Total
Gross Count:	870.00	100.00
Net Area:	865.80	99.52
Background:	4.20	0.48
Composite Fit:	838.00	96.32
Residuals:	32.00	3.68


Analyzed By: Sm

Checked By: SD

000024

042417

AlphaVision Relative Region-Of-Interest (Slope Recalibration)



Acquired: 08:01:23 on 19-May-2004
File: C:\User\Alpha\ALPHA\O42417.SPC
Sample: 0404241-7 PL040513-1

Real Time: 18000.48 s. Live Time: 18000.00 s.
Detector: #4 MCB 1 Input 4
Type: PO-210

520000

Paragon Analytics

Alpha Spectroscopy Analysis

Report Printed:
5/19/04 4:45:06 PM

Para0327.rpt
rev 11/13/03 KVG

Sample Name: 0404241-7D PL040513-1

Analysis Type: PO-210

Detector: MCB 1 Input 5

Date/Time of Count: 5/19/04 8:01:49 AM

Sample Volume: 0.519 Total, 0.260 Aliquot.

Live Time: 300.00 Minutes

Chem. Yield: 73.16%

Real Time: 300.01 Minutes

Total Eff.: 23.41 %

Dead Time: 0.0 %

Tracer Amount: 18.666 DPM.

Acquisition: 512 Channels

Efficiency: 31.99%

Analysis: Relative Region-Of-Interest

Original: 2,965 + 10.3501 * Chn + -0.00110 * Chn **2.

Spectrum Calibration: 2,965 + 10.3699 * Chn + -0.00110 * Chn **2.

Cal File:

Spectrum File: C:\User\Alpha\ALPHA\O42417D.SPC

Background File: C:\USER\ALPHA\BKGND\B4051805.SPC

Library File: C:\User\Alpha\ALPHAVIS.ALB

Peaks

Peak	Channel	Start	End	FWHM	Height	Gross Cts	Bkg Cts	Net Area	DPM
1	231.25	211	236	2.00	6.00	30.00	0.60	29.40	0.42
Tracer	188.62	168	194	4.00	122.00	657.00	1.80	655.20	9.33

Analysis Results

Peak	Nuclide	Energy (keV)	Width (keV)	Aliquot pCi	MDA pCi	% Error
1	Po-210	5304.38	19.72	0.727	n/a	36.55 %
Tracer	Po-209	4882.00	39.82	16.191	n/a	7.65 %

Totals

		% Total
Gross Count:	730.00	100.00
Net Area:	724.00	99.18
Background:	6.00	0.82
Composite Fit:	687.00	94.11
Residuals:	43.00	5.89


Analyzed By: *Sm*

Checked By: *SD*

000026

042417D

AlphaVision Relative Region-Of-Interest (Slope Recalibration)



Acquired: 08:01:49 on 19-May-2004
File: C:\User\Alpha\ALPHA\O42417D.SPC
Sample: 0404241-7D PL040513-1

Real Time: 18000.48 s. Live Time: 18000.00 s.
Detector: #5 MCB 1 Input 5
Type: PO-210

20000

Paragon Analytics

Alpha Spectroscopy Analysis

Report Printed:
5/19/04 4:47:13 PM

Para0327.rpt
rev 11/13/03 KVG

Sample Name: 0404241-9 PL040513-1

Analysis Type: PO-210

Detector: MCB 1 Input 7

Date/Time of Count: 5/19/04 8:02:20 AM

Sample Volume: 0.506 Total, 0.253 Aliquot.

Live Time: **300.00 Minutes**

Chem. Yield: 85.62%

Real Time: 300.01 Minutes

Total Eff.: 27.23 %

Dead Time: 0.0 %

Tracer Amount: 18.666 DPM.

Acquisition: 512 Channels

Efficiency: 31.80%

Analysis: Relative Region-Of-Interest

Original: 2,967 + 10.3929 * Chn + -0.00120 * Chn **2.

Spectrum Calibration: 2,967 + 10.4944 * Chn + -0.00120 * Chn **2.

Cal File:

Spectrum File: C:\User\Alpha\ALPHA\O42419.SPC

Background File: C:\USER\ALPHA\BKGND\B4051807.SPC

Library File: C:\User\Alpha\ALPHAVIS.ALB

Peaks

Peak	Channel	Start	End	FWHM	Height	Gross Cts	Bkg Cts	Net Area	DPM
1	228.72	210	236	6.00	660.00	3,919.00	0.00	3919.00	47.98
Tracer	186.47	168	193	6.00	127.00	764.00	1.50	762.50	9.33

Analysis Results

Peak	Nuclide	Energy (keV)	Width (keV)	Aliquot pCi	MDA pCi	% Error
1	Po-210	5304.38	59.67	85.422	n/a	3.13 %
Tracer	Po-209	4882.00	60.28	16.620	n/a	7.09 %

Totals

		% Total
Gross Count:	4,741.00	100.00
Net Area:	4,736.80	99.91
Background:	4.20	0.09
Composite Fit:	4,683.00	98.78
Residuals:	58.00	1.22


Analyzed By: SM

Checked By: SD

000028

042419

AlphaVision Relative Region-Of-Interest (Slope Recalibration)



Real Time: 18000.48 s. Live Time: 18000.00 s.
Detector: #7 MCB 1 Input 7
Type: PO-210

Acquired: 08:02:20 on 19-May-2004
File: C:\User\Alpha\ALPHA\O42419.SPC
Sample: 0404241-9 PL040513-1

620000

Paragon Analytics

Alpha Spectroscopy Analysis

Report Printed:
5/19/04 4:48:43 PM

Para0327.rpt
rev 11/13/03 KVG

Sample Name: 0404241-11 PL040513-1

Analysis Type: PO-210

Detector: MCB 1 Input 8

Date/Time of Count: 5/19/04 8:02:41 AM

Sample Volume: 0.503 Total, 0.251 Aliquot.

Live Time: 300.00 Minutes

Chem. Yield: 83.05%

Real Time: 300.01 Minutes

Total Eff.: 26.23 %

Dead Time: 0.0 %

Tracer Amount: 18.666 DPM.

Acquisition: 512 Channels

Efficiency: 31.59%

Analysis: Relative Region-Of-Interest

Original: 2,960 + 10.3594 * Chn + -0.00120 * Chn **2.

Spectrum Calibration: 2,960 + 10.4752 * Chn + -0.00120 * Chn **2.

Cal File:

Spectrum File: C:\User\Alpha\ALPHA\O424111.SPC

Background File: C:\USER\ALPHA\BKGND\B4051808.SPC

Library File: C:\User\Alpha\ALPHAVIS.ALB

Peaks

Peak	Channel	Start	End	FWHM	Height	Gross Cts	Bkg Cts	Net Area	DPM
1	229.85	210	237	8.00	144.00	932.00	0.30	931.70	11.84
Tracer	187.51	168	193	4.00	111.00	735.00	0.60	734.40	9.33

Analysis Results

Peak	Nuclide	Energy (keV)	Width (keV)	Aliquot pCi	MDA pCi	% Error
1	Po-210	5304.38	79.39	21.228	n/a	6.42 %
Tracer	Po-209	4882.00	40.10	16.733	n/a	7.23 %

Totals

		% Total
Gross Count:	1,709.00	100.00
Net Area:	1,706.30	99.84
Background:	2.70	0.16
Composite Fit:	1,667.00	97.54
Residuals:	42.00	2.46


Analyzed By: Sm

Checked By: SD

000030

O424111

AlphaVision Relative Region-Of-Interest (Slope Recalibration)



Acquired: 08:02:41 on 19-May-2004

File: C:\User\Alpha\ALPHA\O424111.SPC

Sample: 0404241-11 PL040513-1

Real Time: 18000.48 s. Live Time: 18000.00 s.


Detector: #8 MCB 1 Input 8

Type: PO-210

130000

0424112

AlphaVision Relative Region-Of-Interest (Slope Recalibration)



Acquired: 08:03:07 on 19-May-2004

File: C:\User\Alpha\ALPHA\O424112.SPC

Sample: 0404241-12 PL040513-1

Real Time: 18000.16 s. Live Time: 18000.00 s.

Detector: #49 MCB 7 Input 1

Type: PO-210

000003

Paragon Analytics

Alpha Spectroscopy Analysis

Report Printed:
5/19/04 4:50:35 PM

Para0327.rpt
rev 11/13/03 KVG

Sample Name: PL040513-1MB PL040513-1

Analysis Type: PO-210

Detector: MCB 7 Input 2

Date/Time of Count: 5/19/04 8:03:38 AM

Sample Volume: 0.500 Total, 0.250 Aliquot.

Live Time: **300.00 Minutes**

Chem. Yield: 76.39%

Real Time: 300.00 Minutes

Total Eff.: 23.99 %

Dead Time: 0.0 %

Tracer Amount: 18.666 DPM.

Acquisition: 512 Channels

Efficiency: 31.41%

Analysis: Relative Region-Of-Interest

Original: 3,024 + 9.9726 * Chn + -0.00036 * Chn **2.

Spectrum Calibration: 3,024 + 10.0244 * Chn + -0.00036 * Chn **2.

Cal File:

Spectrum File: C:\User\Alpha\ALPHA\O40511B.SPC

Background File: C:\USER\ALPHA\BKGND\B4051850.SPC

Library File: C:\User\Alpha\ALPHAVIS.ALB

Peaks

Peak	Channel	Start	End	FWHM	Height	Gross Cts	Bkg Cts	Net Area	DPM
1	229.36	209	234	2.00	1.00	1.00	0.00	1.00	0.01
Tracer	186.59	166	192	4.00	84.00	673.00	1.20	671.80	9.33

Analysis Results

Peak	Nuclide	Energy (keV)	Width (keV)	Aliquot pCi	MDA pCi	% Error
1	Po-210	5304.38	19.72	0.025	n/a	196.00 %
Tracer	Po-209	4882.00	39.56	16.816	n/a	7.56 %

Totals

	Totals	% Total
Gross Count:	741.00	100.00
Net Area:	735.00	99.19
Background:	6.00	0.81
Composite Fit:	674.00	90.96
Residuals:	67.00	9.04


Analyzed By: Sm

Checked By: SD

000034

O40511B

AlphaVision Relative Region-Of-Interest (Slope Recalibration)



Acquired: 08:03:38 on 19-May-2004

File: C:\User\Alpha\ALPHA\O40511B.SPC

Sample: PL040513-1MB PL040513-1

Real Time: 18000.18 s. Live Time: 18000.00 s.

Detector: #50 MCB 7 Input 2

Type: PO-210

000035

Paragon Analytics

Alpha Spectroscopy Analysis

Report Printed:
5/21/04 9:34:06 AM

Para0327.rpt
rev 11/13/03 KVG

Sample Name: PL040513-1LCS PL040513-1

Analysis Type: PO-210

Detector: MCB 7 Input 3

Date/Time of Count: 5/19/04 8:03:58 AM

Sample Volume: 0.500 Total, 0.250 Aliquot.

Live Time: **300.00 Minutes**

Chem. Yield: 67.58%

Real Time: 300.00 Minutes

Total Eff.: 20.91 %

Dead Time: 0.0 %

Tracer Amount: 18.666 DPM.

Acquisition: 512 Channels

Efficiency: 30.94%

Analysis: Relative Region-Of-Interest

Original: 2,999 + 10.1950 * Chn + -0.00086 * Chn **2.

Spectrum Calibration: 2,999 + 10.2707 * Chn + -0.00086 * Chn **2.

Cal File:

Spectrum File: C:\User\Alpha\ALPHA\O40511L.SPC

Background File: C:\USER\ALPHA\BKGND\B4051851.SPC

Library File: C:\User\Alpha\ALPHAVIS.ALB

Peaks

Peak	Channel	Start	End	FWHM	Height	Gross Cts	Bkg Cts	Net Area	DPM
1	228.80	212	234	6.00	203.00	1,113.00	0.30	1112.70	17.74
Tracer	186.20	165	192	6.00	82.00	586.00	0.60	585.40	9.33

Analysis Results

Peak	Nuclide	Energy (keV)	Width (keV)	Aliquot pCi	MDA pCi	% Error
1	Po-210	5304.38	59.27	31.963	n/a	5.88 %
Tracer	Po-209	4882.00	59.71	16.816	n/a	8.10 %

Totals

% Total

Gross Count:	1,842.00	100.00
Net Area:	1,836.30	99.69
Background:	5.70	0.31
Composite Fit:	1,699.00	92.24
Residuals:	143.00	7.76


Analyzed By: JP

Checked By: SD

000036

O40511L

AlphaVision Relative Region-Of-Interest (Slope Recalibration)



Acquired: 08:03:58 on 19-May-2004

File: C:\User\Alpha\ALPHA\O40511L.SPC

Sample: PL040513-1LCS PL040513-1

Real Time: 18000.16 s. Live Time: 18000.00 s.

Detector: #51 MCB 7 Input 3

Type: PO-210

730000

Paragon Analytcs
Alpha Spectrometer Instrument Run Log

Date: 5/19/04

277729

Detector	Batch ID	Sample ID	Iso/Matrix	Duration	Initial	File ID/Comm.
9	AS040510-3	AS040510-31CS	V/W	300	JP	VR5103LD
10	AS040510-4	0404248-8	T/W	300	JP	TR42488
11		0404313-1				T43131
12		↓ -2				↓ 2
13		0404314-1				T43141
15		↓ -4				↓ 4
16		↓ -5				↓ 5
17		↓ -8				↓ 8
18		↓ -23				↓ 23
19		↓ -24				↓ 24
21		AS040510-4MB		420		T5104B
22		↓ LOS		300		↓ L
23		↓ LOS				↓ LD
24	AS040511-4	0404228-1	V/S	300	JP	U42281
42		↓ -2				↓ 2
43		↓ -3				↓ 3
44		↓ -4				↓ 4
45		↓ -5				↓ 5
46		↓ -6				↓ 6
47		↓ -7				↓ 7
48		↓ -8				↓ 8
57		↓ -8D				↓ 8D
58		↓ -9				↓ 9
59		↓ -10				↓ 10

Detector	Batch ID	Sample ID	Iso/Matrix	Duration	Initial	File ID/Comm.
60	AS040511-4	0404228-11	V/S	300	JP	U422811
61		↓ -12				↓ 12
62		↓ -13				↓ 13
63		↓ -14				↓ 14
64		↓ -15				↓ 15
2	PI040513-1	0404241-3	Pe/S	300	JP	042413
3		↓ -5				↓ 5
4		↓ -7				↓ 7
5		↓ -7D				↓ 7D
7		↓ -9				↓ 9
8		↓ -11				↓ 11
49		↓ -12				↓ 12
50		PI040513-1MB				05131B
51		↓ LOS				↓ L
9	AS040511-4	0404228-16	U/S	300	Am	U422816
10		↓ -17				↓ 17
11		↓ -18				↓ 18
12		↓ -18D				↓ 18D
13		↓ -19				↓ 19
15		↓ -20				↓ 20
16		↓ -4MB				U5114B
17		↓ -4LS				U5114L
18	AS040428-5	0404241-9	U/S	300	Am	U42419
19		↓ -12				U424112

Notes:

Cont. on page 277730. JP 5/19/04

Reviewed by: JD
Date: 5/19/04

PARAGON ANALYTICS
Radiochemistry Data Package

Section 5

**QUALITY ASSURANCE
SUMMARY REPORTS**

5

No *NON-CONFORMANCE REPORTS* or
QUALITY ASSURANCE SUMMARY SHEETS
are included in this data package.

~~000033~~
000040

PARAGON ANALYTICS
Radiochemistry Data Package

Section 6

**LABORATORY
BENCH SHEETS**

6

000041

Radiochemistry Instrument Worksheet

Prep Batch: PL040513-1

Prep Batch: PL040513-1

Prep Procedure: Po210

Analytical QASS / NCR? Y (N) NA

Prep Num	LabID	QC Type	Init Alq	Fin Alq	Units	Cnt 1 File	Cnt 1 Inst/Det	Cnt 1 Pos Chk By	Cnt 2 File	Cnt 2 Inst/Det	Cnt 2 Pos Chk By	Cnt 3 File	Cnt 3 Inst/Det	Cnt 3 Pos Chk By	Notes
1	0404241-3	SMP	0.534	0.267	g	0_42413	2	<u>SM</u>	_42413			_42413			
1	0404241-5	SMP	0.5106	0.2553	g	_42415	3		_42415			_42415			
1	0404241-7	SMP	0.51	0.255	g	_42417	4		_42417			_42417			
1	0404241-7	DUP	0.5193	0.25965	g	_42417D	5		_42417D			_42417D			
1	0404241-9	SMP	0.5059	0.25295	g	_42419	7		_42419			_42419			
1	0404241-11	SMP	0.5025	0.25125	g	_424111	8		_424111			_424111			
1	0404241-12	SMP	0.5018	0.2509	g	_424112	49		_424112			_424112			
1	PL040513-1	MB	0.5	0.25	g	_40511B	50		_40511B			_40511B			
1	PL040513-1	LCS	0.5	0.25	g	_40511L	51		_40511L			_40511L			

JP 5/21/04

Tracer/Carrier Solution Information

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
T1	Po-209	592.1808.81	37.331	DPM/ml	05/13/04	0.5	ml	AW-013

Spike Solution Information

Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date	Aliquot	Units	Pipet ID
S1	Pb-210	059.1572.15	32.153	DPM/ml	05/13/04	1	ml	AW-010
S1	Po-210	059.1572.15	32.153	DPM/ml	05/13/04	1	ml	AW-010

0000

Radiochemistry Prep Worksheet

Prep Batch: PL040513-1

Prep Procedure: Po210

Reviewed By: CRW *CRW* Review Date: 5/18/04

Non-Routine Pre-Treatment? Y N Batch: *N/A* Re-Prep? Y N Prep QASS / NCR? Y N

Prep SOP: PAI 711 Rev: 5
 Prep SOP: NONE
 Matrix Class: solid

Prep Analyst: Chad Wangeline
 Prep Date: 5/13/04
 Prep Dept: AP

Balance: 23
 Balance:

Sample Num	Prep Num	LabID	QC Type	Dish No.	Init Aliq g	Fin Aliq g	Prep Basis	Decay Date/Time	Standards	Prep Notes
1	1	0404241-3	SMP		0.534	0.267	Dry Weight	05/15/04 10:00	T1	<i>CRW 5/18/04</i>
2	1	0404241-5	SMP		0.5106	0.2553	Dry Weight	05/15/04 10:00	T1	
3	1	0404241-7	SMP		0.51	0.255	Dry Weight	05/15/04 10:00	T1	
4	1	0404241-7	DUP		0.5193	0.25965	Dry Weight	05/15/04 10:00	T1	
5	1	0404241-9	SMP		0.5059	0.25295	Dry Weight	05/15/04 10:00	T1	
6	1	0404241-11	SMP		0.5025	0.25125	Dry Weight	05/15/04 10:00	T1	
7	1	0404241-12	SMP		0.5018	0.2509	Dry Weight	05/15/04 10:00	T1	
8	1	PL040513-1	MB		0.5	0.25	Dry Weight	05/15/04 10:00	T1	
9	1	PL040513-1	LCS		0.5	0.25	Dry Weight	05/15/04 10:00	S1, T1	

Spiked By: Chad Wangeline Date: 5/13/04
 Witnessed By: Grace Campagnola Date: 5/13/04

Relinquished By: *CRW* Date: 5/18/04
 Received By: *SPD* Date: 5/19/04

Tracer/Carrier Solution Information					
Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date
T1	Po-209	592.1808.81	37.331	DPM/ml	05/13/04
				0.5	ml
					AW-013

Spike Solution Information					
Soln #	Nuclide	SolnID	Prep Conc	Units	Prep Date
S1	Pb-210	059.1572.15	32.153	DPM/ml	05/13/04
				1	ml
					AW-010
S1	Po-210	059.1572.15	32.153	DPM/ml	05/13/04
				1	ml
					AW-010

0000

Comments: ALL SAMPLES COVERED WITH THIN FILM.

Radiochemistry Prep Worksheet

Paragon Analytics

Prep Batch: PL040513-1

Prep Procedure: Po210

Prep Batch Not Validated!!!

Reviewed By:

Review Date:

Non-Routine Pre-Treatment? Y / N Batch: _____

Re-Prep? Y / N Batch: _____

Prep QASS / NCR? Y / N _____

Prep SOP: PAI 711 Rev: 5

Prep Analyst: Chad Wangeline *CW*

Balance:

Prep Date: 5/13/04

Balance:

Prep Dept: AP

Matrix Class: solid

Sampl Num	Prep Num	LabID	QC Type	Dish No.	Init Alq g	Fin Alq g	Prep Basis	Decay Date/Time	Prep Notes
1	1	0404241-3	SMP		0.534	0.267	Dry Weight	5/15/04 1000	TI
2	1	0404241-5	SMP		0.5106	0.2553	Dry Weight		TI
3	1	0404241-7	SMP		0.51	0.255	Dry Weight		TI
4	1	0404241-7	DUP		0.5193	0.25965	Dry Weight		TI
5	1	0404241-8	SMP		0.5059	0.25285	Dry Weight		TI
6	1	0404241-11	SMP		0.5025	0.25125	Dry Weight		TI
7	1	0404241-12	SMP		0.5018	0.2509	Dry Weight		TI
8	1	PL040513-1	MB		0.5	0.25			TI, SI
9	1	PL040513-1	LCS		0.5	0.25			TI, SI

Spiked By: *CW* Date: 5/13/04

Witnessed By: *CW* Date: 5/13/04

Relinquished By:

Date:

Received By:

Date:

TI P6-209 592.1808.81 0.5 Awo13 SI P6-210 059.1572.15 1.000 Awo10
exp 2/10/05 exp 11/7/04

Comments

SAMPLE CONDITION FORM (SOLIDS)

ANALYST: CW

ANALYSIS DATE: 5/13/04

METHOD: P-210

WORK ORDER	SAMPLE ID	SAMPLE CONDITION		
		Dry/Wet	TEXTURE	Remarks
0404241	3	Dry	Ground	None
↓	5	↓	↓	↓
↓	7	↓	↓	↓
↓	9	↓	↓	↓
↓	11	↓	↓	↓
↓	12	↓	↓	↓
<div style="position: relative; height: 100px;"> CW 5/13/04 </div>				

4/28/04

W.O. # 04-04-241 / (PD^{ISO}, Th^{ISO}, U^{ISO}, Pb 210L)

in oven - 4/27 @ 1100

out oven - 4/28 @ 1100

S.O.P. - 721-R10

balance # 15, oven # 1
5 balls, shake 15 min.

<u>Sample #</u>	<u>sample / can/balls(g)</u>	<u>can/balls(g)</u>	<u>ground sample(g)</u>	<u>TYPE sample</u>
04-04-241-3	126.6	96.4	30.2	SOIL
-5	126.5	96.3	30.2	
-7	126.8	96.4	30.4	
-9	127.0	96.9	30.1	
-11	127.1	96.7	30.4	
-12	125.7	96.5	29.2	

C.O. - 4/28/04

Continued on Page _____

Read and Understood By

Carson B. O'Dell 4/28/04

Signed

Date

Trace A

Signed

4/28/04 000046

Date

PARAGON ANALYTICS
Radiochemistry Data Package

Section 7

**STANDARDS
TRACEABILITY
DOCUMENTS**

7

000047

Po-209 Spiking Solution

1) Determine density of 2.0 N HCl
 Mass of 100 ml volumetric flask → 56.4602 g Bal 12
 Mass of flask + HCl → 159.5771 g ↓
 Net mass of HCl → 103.1169 g
 $\rho = 1.031 \text{ g/ml}$

2) Transfer contents of RSO# 592 ampoule to a 1 l wide-mouth poly bottle

Mass of 1 l bottle (w/o lid) 72.2002 g Bal 12
 Mass of ampoule (opened + 50 ml beaker) 327.0628 g Bal 12
 Mass of ampoule, beaker w/o std 32.0264 g ↓
 Net mass of std. transferred 5.0364 g

3) Add 217 HCl to final dilution

Mass of 1 l bottle (from above) 72.2002 g Bal 12
 Mass of bottle + std + 217 HCl 737.5 g Bal 26
 Net mass of std. 665.3 g

4) Final Activity Calculation:

$$\frac{(85.42 \text{ Bq/g}) \times (60 \text{ dpm/Bq}) \times (5.0364 \text{ g}) \times (1.031 \text{ g/ml})}{(665.3 \text{ g})} = 40.00 \text{ dpm/ml}$$

Std ID: 592.1808.81

Description: Po-209

Activity: 40.00 dpm/ml

Uncertainty: 0.168 dpm/ml

Ref. Date: 3/15/94

Ref Time: na

Prep Date: 11/15/01 Prep by: DCB

Expiration: 11/15/02

Matrix/Comp. 2.0N HCl

Half Life (y): 1.02E+02

Standard Reverified on 11-17-02.

Expires: 11-17-03

RG 12/6/01

Standard Reverified on 2-10-04.

Expires: 2-10-05

RG 2/17/04

Continued on Page

Read and Understood By

Dianna Burn

11/15/01

Dance Kelley

12/6/01

Signed

Date

Signed

Date

000048



National Institute of Standards & Technology

Certificate

PA ID 0592
recd 9-25-01

Standard Reference Material 4326 Polonium-209 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive polonium-209 chloride and hydrochloric acid dissolved in 5 mL of distilled water. The solution is contained in a flame-sealed NIST borosilicate-glass ampoule. The SRM is intended for the calibration of alpha-particle counting instruments and for the monitoring of radiochemical procedures.

Radiological Hazard

The SRM ampoule contains polonium-209 with a total activity of approximately 500 Bq. Polonium-209 decays by alpha-particle emission. None of the alpha particles escape from the SRM ampoule. During the decay process X-rays and gamma rays with energies from 10 to 900 keV are also emitted. Most of these photons escape from the SRM ampoule but their intensities are so small that they do not represent a radiation hazard. Approximate unshielded dose rates at several distances (as of the reference time) are given in note [a]*. The SRM should be used only by persons qualified to handle radioactive material.

Chemical Hazard

The SRM ampoule contains hydrochloric acid (HCl) with a concentration of 2 moles per liter of water. The solution is corrosive and represents a health hazard if it comes in contact with eyes or skin. If the ampoule is to be opened to transfer the solution, the recommended procedure is given on page 2. The ampoule should be opened only by persons qualified to handle both radioactive material and strong acid solution.

Storage and Handling

The SRM should be stored and used at a temperature between 5 and 65 °C. The solution in an unopened ampoule should remain stable and homogeneous until at least June 2004. Refer to reference [5] for details on the long-term stability of polonium solution standards.

The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material. If the ampoule is transported it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) both because of the radioactivity and because of the strong acid.

Preparation

This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, J.M.R. Hutchinson, Group Leader. The overall technical direction and physical measurements leading to certification were provided by R. Collé of the Radioactivity Group and Z. Lin, Guest Researcher.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by N.M. Trahey.

Gaithersburg, Maryland 20899
January 1995 (Text only revised November 1997)

Thomas E. Gills, Chief
Standard Reference Materials Program

Liquid-Scintillation Counting Warning

Polonium-209 decays primarily by alpha-particle emission. One of the principal alpha-particle transitions feeds a low-energy, **delayed**, isomeric state in the lead-205 daughter. Liquid-scintillation measurements of polonium-209 activity can include some of the activity of this isomeric state. Refer to reference [6] for further information about the effect of this isomeric transition on routine liquid-scintillation counting of polonium-209.

Recommended Procedure for Opening the SRM Ampoule

- 1) If the SRM solution is to be diluted, it is recommended that the diluting solution have an acid concentration comparable to that of the SRM solution.
- 2) Wear eye protection, gloves, and protective clothing and work over a tray with absorbent paper in it. Work in a fume hood. In addition to the radioactive material, the solution contains strong acid and is corrosive.
- 3) Shake the ampoule to wet all of the inside surface of the ampoule. Return the ampoule to the upright position.
- 4) Check that all of the liquid has drained out of the neck of the ampoule. If necessary, gently tap the neck to speed the process.
- 5) Holding the ampoule upright, score the narrowest part of the neck with a scribe or diamond pencil.
- 6) Lightly wet the scored line. This reduces the crack propagation velocity and makes for a cleaner break.
- 7) Hold the ampoule upright with a paper towel, a wiper, or a support jig. Position the scored line away from you. Using a paper towel or wiper to avoid contamination, snap off the top of the ampoule by pressing the narrowest part of the neck away from you while pulling the tip of the ampoule towards you.
- 8) Transfer the solution from the ampoule using a pycnometer or a pipet with dispenser handle. **NEVER PIPETTE BY MOUTH.**
- 9) Seal any unused SRM solution in a flame-sealed glass ampoule, if possible, to minimize the evaporation loss.

See also reference [4]*.

PROPERTIES OF SRM 4326
(Certified values are shown in bold type)

Source identification number	NIST SRM 4326		
Physical Properties:			
Source description	Liquid in flame-sealed NIST borosilicate-glass ampoule		
Ampoule specifications	Body outside diameter	(16.5 ± 0.5) mm	
	Wall Thickness	(0.60 ± 0.04) mm	
	Barium content	Less than 2.5%	
	Lead-oxide content	Less than 0.02%	
	Other heavy elements	Trace quantities	
Solution density	(1.031 ± 0.004) g·mL ⁻¹ at 22 °C [b]*		
Solution mass	(5.160 ± 0.003) g [b]		
Chemical Properties:			
Solution composition	Chemical Formula	Concentration (mol·L ⁻¹)	Mass Fraction (g·g ⁻¹)
	H ₂ O	53	0.93
	HCl	2	0.07
	HNO ₃	<3 × 10 ⁻³	<2 × 10 ⁻⁴
	PoCl ₄	7 × 10 ⁻¹⁰	2 × 10 ⁻¹⁰
Radiological Properties:			
Radionuclide	Polonium-209		
Reference time	1200 EST, 15 March 1994		
Massic alpha-particle-emission rate of the solution [c]	85.42 α·s ⁻¹ ·g ⁻¹ (Polonium-209 only) [d]		
Relative expanded uncertainty (k=2)	0.42% [e]		
Alpha-particle-emitting impurities	Polonium-208: (0.106 ± 0.017) α·s ⁻¹ ·g ⁻¹ [b, f]		
Photon-emitting impurities	None detected [g]		
Half lives used	Polonium-209: (102 ± 5) a [h] Polonium-208: (2.898 ± 0.002) a [h]		
Measuring instruments	Two 4π α liquid-scintillation counting systems, 2π α gas-flow proportional counter, and silicon surface-barrier detector		

EVALUATION OF THE UNCERTAINTY OF THE MASSIC ALPHA-PARTICLE-EMISSION RATE
[e]*

Input Quantity x_i , the source of uncertainty (and individual uncertainty components where appropriate)	Method Used To Evaluate $u(x_i)$, the standard uncertainty of x_i (A) denotes evaluation by statistical methods (B) denotes evaluation by other methods	Relative Uncertainty Of Input Quantity, $u(x_i)/x_i$, (%) [i]	Relative Sensitivity Factor, $ \partial y/\partial x_i \cdot$ (x_i/y) [j]	Relative Uncertainty Of Output Quantity, $u_i(y)/y$, (%) [k]
Massic liquid-scintillation count rate, corrected for background and decay	Standard deviation for repeated measurements. Six degrees of freedom. (A)	0.06	1.0	0.06
Background variability	Multiple comparisons (A) [7]	0.20	0.02 [m]	0.004
Liquid-scintillator quench corrections	Multiple comparisons (A) [7]	0.12	1.0	0.12
Liquid-scintillation- cocktail stability	Multiple comparisons (A) [7]	0.7	0.007	0.005
Gravimetric measurements	Estimated (B)	0.05	1.0	0.05
Half life of Po-208 Half life of Po-209	Standard uncertainty of the half life (A)	0.07 [n] 4.9 [n]	0.0003 [p] 0.004 [p]	0.00002 0.02
Extrapolation of alpha- particle-count-rate- versus-energy to zero energy	Estimated (B) [7]	0.06	1.0	0.06
Live-time [q]	Estimated (B) [7]	0.04	1.0	0.04
Alpha-particle detection efficiency of scintillator	Estimated (B) [7]	0.10	1.0	0.10
Correction for non- alpha-particle decay modes	Estimated (B) [7]	0.06	1.0	0.06
Alpha-particle-emitting impurities	Estimated (B) [r] Limit of detection (B) [s]	8.1 100.	0.001 0.0006	0.01 0.06
Photon-emitting impurities	Limit of detection (B) [s]	100.	0.0002	0.02
Relative Combined Standard Uncertainty of the Output Quantity, $u_c(y)/y$, (%)				0.21
Coverage Factor, k				<u>x 2</u>
Relative Expanded Uncertainty of the Output Quantity, U/y , (%)				0.42

NOTES

- [a] The Sievert is the SI unit for dose equivalent. See reference [1]. One μSv is equal to 0.1 mrem.
 Distance from Ampoule (cm): 1 30 100
 Approximate Dose Rate ($\mu\text{Sv/h}$): <0.1 - -
- [b] The stated uncertainty is two times the standard uncertainty.
- [c] **Massic alpha-particle-emission rate** is the preferred name for the quantity alpha-particle-emission rate divided by the total mass of the sample. **Massic activity** is the preferred name for the quantity activity divided by the total mass of the sample. See reference [1].
- [d] The polonium-209 massic activity of the solution is $85.83 \text{ Bq}\cdot\text{g}^{-1}$, assuming an alpha-particle branching ratio of $(0.9952 \pm 0.0004) \alpha\cdot\text{s}^{-1}\cdot\text{Bq}^{-1}$ [h].
- [e] The reported value, y , of massic alpha-particle-emission rate (alpha-particle-emission rate per unit mass) at the reference time was not measured directly but was derived from measurements and calculations of other quantities. This can be expressed as $y = f(x_1, x_2, x_3, \dots, x_n)$, where f is a mathematical function derived from the assumed model of the measurement process.
- The value, x_i , used for each input quantity i has a **standard uncertainty**, $u(x_i)$, that generates a corresponding uncertainty in y , $u_i(y) \equiv |\partial y/\partial x_i| \cdot u(x_i)$, called a **component of combined standard uncertainty** of y .
- The **combined standard uncertainty** of y , $u_c(y)$, is the positive square root of the sum of the squares of the components of combined standard uncertainty.
- The combined standard uncertainty is multiplied by a **coverage factor** of $k = 2$ to obtain U , the **expanded uncertainty** of y .
- Since it can be assumed that the possible estimated values of the massic alpha-particle-emission rate are approximately normally distributed with approximate standard deviation $u_c(y)$, the unknown value of the massic alpha-particle-emission rate is believed to lie in the interval $y \pm U$ with a level of confidence of approximately 95 percent.
- For further information on the expression of uncertainties, see references [2] and [3]. See reference [7] for further information on uncertainties related to this calibration.
- [f] Estimated limits of detection for alpha-particle-emitting impurities are:
 0.002 $\alpha\cdot\text{s}^{-1}\cdot\text{g}^{-1}$ for energies less than 3.5 MeV,
 0.05 $\alpha\cdot\text{s}^{-1}\cdot\text{g}^{-1}$ for energies between 3.5 and 4.2 MeV, and
 0.0002 $\alpha\cdot\text{s}^{-1}\cdot\text{g}^{-1}$ for energies greater than 5.18 MeV.
- [g] Estimated limits of detection for photon-emitting impurities are:
 $2 \times 10^{-4} \gamma\cdot\text{s}^{-1}\cdot\text{g}^{-1}$ for energies between 15 and 68 keV,
 $2 \times 10^{-4} \gamma\cdot\text{s}^{-1}\cdot\text{g}^{-1}$ for energies between 81 and 256 keV,
 $6 \times 10^{-5} \gamma\cdot\text{s}^{-1}\cdot\text{g}^{-1}$ for energies between 266 and 892 keV, and
 $4 \times 10^{-6} \gamma\cdot\text{s}^{-1}\cdot\text{g}^{-1}$ for energies between 900 and 3300 keV,
 provided that the photons are separated in energy by 4 keV or more from photons emitted in the decay of polonium-209. See reference [7] for further information about the impurity analyses.
- [h] The stated uncertainty is the standard uncertainty. See reference [8].

- [i] Relative standard uncertainty of the input quantity x_i .
- [j] The relative change in the output quantity y divided by the relative change in the input quantity x_i . If $|\partial y/\partial x_i| \cdot (x_i/y) = 1.0$, then a 1% change in x_i results in a 1% change in y . If $|\partial y/\partial x_i| \cdot (x_i/y) = 0.05$, then a 1% change in x_i results in a 0.05% change in y .
- [k] Relative component of combined standard uncertainty of output quantity y , rounded to two significant figures or less. The relative component of combined standard uncertainty of y is given by $u_i(y)/y \equiv |\partial y/\partial x_i| \cdot u(x_i)/y = |\partial y/\partial x_i| \cdot (x_i/y) \cdot u(x_i)/x_i$. The numerical values of $u(x_i)/x_i$, $|\partial y/\partial x_i| \cdot (x_i/y)$, and $u_i(y)/y$, all dimensionless quantities, are listed in columns 3, 4, and 5, respectively. Thus, the value in column 5 is equal to the value in column 4 multiplied by the value in column 3. The input quantities are independent, or very nearly so. Hence the covariances are zero or negligible.
- [m] $|\partial y/\partial x_i| \cdot (x_i/y) = (\text{average background count rate})/(\text{average net sample count rate})$.
- [n] The relative standard uncertainty of $\lambda \cdot t$ is determined by the relative standard uncertainty of λ (i.e., of the half life). The relative standard uncertainty of t is negligible.
- [p] $|\partial y/\partial x_i| \cdot (x_i/y) = |\lambda \cdot t| \cdot \{(\text{response per Bq of impurity})/(\text{response per Bq of Po-209})\} \cdot \{(\text{Bq of impurity})/(\text{Bq of Po-209})\}$.
- [q] The live time is determined by counting the pulses from a gated oscillator.
- [r] The standard uncertainty given is for the detected Po-208 impurity. $|\partial y/\partial x_i| \cdot (x_i/y) = \{(\text{response per Bq of impurity})/(\text{response per Bq of Po-209})\} \cdot \{(\text{Bq of impurity})/(\text{Bq of Po-209})\}$.
- [s] The standard uncertainty for each undetected impurity that might reasonably be expected to be present is estimated to be equal to the estimated limit of detection for that impurity, i.e. $u(x_i)/x_i = 100\%$. $|\partial y/\partial x_i| \cdot (x_i/y) = \{(\text{response per Bq of impurity})/(\text{response per Bq of Po-209})\} \cdot \{(\text{Bq of impurity})/(\text{Bq of Po-209})\}$. Thus $u_i(y)/y$ is the relative change in y if the impurity were present with a massic activity equal to the estimated limit of detection.

REFERENCES

- [1] International Organization for Standardization (ISO), *ISO Standards Handbook - Quantities and Units*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900.
- [2] International Organization for Standardization (ISO), *Guide to the Expression of Uncertainty in Measurement*, 1993. Available from the American National Standards Institute, 11 West 42nd Street, New York, NY 10036, U.S.A. 1-212-642-4900. (Listed under ISO miscellaneous publications as "ISO Guide to the Expression 1993".)
- [3] B. N. Taylor and C. E. Kuyatt, *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*, NIST Technical Note 1297, 1994. Available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20407, U.S.A.
- [4] National Council on Radiation Protection and Measurements Report No. 58, *A Handbook of Radioactivity Measurements Procedures*, Second Edition, 1985. Available from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Bethesda, MD 20814 U.S.A.
- [5] R. Collé, Long-term Stability of Carrier-Free Polonium Solution Standards, *Radioactivity and Radiochemistry*, **4** (1993) 18.
- [6] R. Collé, Z. Lin, J.M.R. Hutchinson, and F.J. Schima, A Delayed Isomeric State in ^{205}Pb and Its Implications for $4\pi\alpha$ Liquid-Scintillation Spectrometry of ^{209}Po , *Appl. Radiat. Isot.*, **45** (1994) 1165.
- [7] R. Collé, Z. Lin, F.J. Schima, P.A. Hodge, J.W.L. Thomas, J.M.R. Hutchinson, and B.M. Coursey, Preparation and Calibration of Carrier-free ^{209}Po Solution Standards, *J. Res. NIST*, **100** (1995) 1.
- [8] Evaluated Nuclear Structure Data File (ENSDF), July 1994.

PROJECT Pb-210 059.1572.15

Continued From Page

8/26/98 Pb-210 standard S-053 is diluted in 1N HCl.

1 ml (1.0654 g) of S-053 was diluted with 1N HCl to total mass of 1018.0 g (1034.6 ml).

Expected Activity = $\frac{1.015 \text{ pCi}}{52.7402 \text{ g}} (1.0654 \text{ g}) (1034.6 \text{ ml})^{-1} = 19.82 \text{ pCi/ml}$

Balance 12

Bottle Tare: 79.2g

Density 1N HCl: 1.0163 g/ml

Bottle Full: 1097.2g
1018.0g

Matrix 1N HCl (BakerLab M 13040)

Label for previous page
 10/6/98
 Strnd ID: 066.1572.14
 Description: Polonium-209 Tracer Solution
 Activity: 20.4 dpm/ml
 Uncertainty: 1.4 dpm/ml
 Ref. Date: 9/22/98
 Ref Time: n/a
 Prep Date: 8/25/98 Prep by: WC
 Expiration: 8/25/01
 Matrix/Comp: 1 M HCl

~~Polonium-209~~
 8/26/98
 Verified
 WC

Strnd ID: 059.1572.15
 Description: Pb-210 working level spiking solution
 Activity: 19.820 pCi/ml
 Uncertainty: 0.991 pCi/ml
 Ref. Date: 4/1/93
 Ref Time: na
 Prep Date: 8/26/98 Prep by: WC
 Expiration: 8/26/03
 Matrix/Comp: 0.1N HCl
 Half Life (y): 2.23E+01

Reverified 11/4/03.
New expiration date 11/7/04.
SD 11/4/03

Continued on Page

Read and Understood By

Wade Cairns

WC

8/8/99

Signed

8/26/98
Date

Signed

Date

000056

PROJECT Pb²¹⁰ stock working^{BS} standard S#053
(4-20-93)

Continued From Page _____

RSD # 00059

Pb-210 std in 2.0M HNO₃

Activity = 1.015 μ Ci in 52.74020 grams

opened ampule 4-20-93 @ 4:00 pm

transferred to plastic bottle (125 ml size)

Purchased from analytics, inc, s/n 45939-387 dy

TAP 5/18/95

Continued on Page _____

Read and Understood By

Robert Jump
Signed

4-20-93
Date

[Signature]
Signed

4-26-93
Date

000057

PROJECT Pb carrier solution

Continued From Page _____

prepared same as p. 3 of this logbook

1.600 g of reagent grade $Pb(NO_3)_2$ dissolved
in 1000 mL of DI H_2O

\therefore 1.0 mL = 1.0 mg Pb

balance used = No. 14

Continued on Page

Read and Understood By

Robert Thompson 9-5-95

000058

Std. 059

CERTIFICATE OF CALIBRATION
Standard Radionuclide Source

45939-307

RECEIVED
APR 13 1993

Pb-210 50 mL Liquid in Flame Sealed Vial

This standard radionuclide source was prepared gravimetrically from a calibrated master liquid radionuclide solution source. The master source was calibrated by liquid scintillation counting.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

Source prepared by: *E. W. Belvin*
E. W. Belvin, Production Manager

ISOTOPE:	Pb-210
ACTIVITY (dps):	3.754 E+04
HALF-LIFE:	22.3 years
CALIBRATION DATE:	April 1, 1993 12:00 EST
TOTAL ERROR:	5.0%
SYSTEMATIC ERROR:	2.6%
RANDOM ERROR:	2.4%

52.74020 grams of solution in 2M HNO₃ containing 0.5 µg Pb and 0.5 µg Bi per mL.

P O NUMBER 31805, Item 1

Q A APPROVED *D. M. Malysz* 4-12-93

PARAGON ANALYTICS
Radiochemistry Data Package

Section 8

CHAIN OF CUSTODY

8

000060

Paragon Analytics

Sample Number(s) Cross-Reference Table

Paragon OrderNum: 0404241

Client Name: New Horizons

Client Project Name: CSMRI

Client Project Number: 2135

Client PO Number:

Client Sample	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
BB04	0404241-1		SOIL	16-Apr-04	13:10
BI17	0404241-2		SOIL	19-Apr-04	11:00
BI17	0404241-3		SOIL	19-Apr-04	11:00
BI37	0404241-4		SOIL	19-Apr-04	14:15
BI37	0404241-5		SOIL	19-Apr-04	14:15
BI42	0404241-6		SOIL	19-Apr-04	15:15
BI42	0404241-7		SOIL	19-Apr-04	15:15
H07	0404241-8		SOIL	22-Apr-04	13:00
H07	0404241-9		SOIL	22-Apr-04	13:00
H07S	0404241-10		SOIL	22-Apr-04	13:05
H07S	0404241-11		SOIL	22-Apr-04	13:05
BB04	0404241-12		SOIL	16-Apr-04	13:10
H07	0404241-13		LEACHAT	22-Apr-04	13:00
H07S	0404241-14		LEACHAT	22-Apr-04	13:05



Paragon Analytics, Inc.

225 Commerce Drive Fort Collins, CO 80524
800-443-1511 or (970) 490-1511 (970) 490-1522 Fax

Accession Number (LAB ID) 0109241

Chain-of-Custody

Date 4/23/04 Page 1 of 1

Sample ID	Date	Time*	Lab ID	Matrix	No. of Containers	Sample(s): <u>see comments</u> (circle one) Turnaround: (Standard or Rush) <u>Standard</u> Dispose or Return to Client	
						Project Name / No.: <u>CSMRI/2135</u>	Project Name / No.: <u>Computer Library + Re-210, B-210</u>
BB04	4/16/04	1310	1	S	1	VOCS	SW8260B E624 E524.2 OLMO
BT17	4/16/04	1100	2/B	S	2	BTEX (only)	SW8021B
BT37	↓	1415	4/5	S	2	SVOCs	SW8270C E625 E525 OLMO
BT42	↓	1515	6/7	S	2	OC Pesticides	SW8081A E608 E508 OLMO
407	4/22/04	1300	8/9	S	2	PCBs	SW8082 E608 E506 OLMO
407S	↓	1305	10/11	S	2	OP Pesticides	SW8141A E614
						Herbicides	SW8151A E615
						TCLP Organics	SW1311 8260B 8270C 8081A 8151A
						TCLP Metals	SW 1311 6010B 7471
						Total Metals	SW6010B 7470 7471 E200 ILMO
						Reactive CN / S	SW846 Chapter 7
						Hexavalent Chromium	SW7196A Alkaline Digest? Y / N
						Inorganic Anions **	SW9056 E300 0
						pH	SW9040B 9045C
						Oil & Grease	SW9071A E413.2
						TPH	GRO DRO SW8015B (both)
						TOX	SW9020B
						Gross Alpha / Beta	SW9310 E900 0
						Actinides by PAI SOP (circle) Pu / U / Am / Th / Cm	
						Total Uranium by KPA	D5174-91
						Tritium	E906.0
						Total Alpha-Emitting Radium	SW9315 E903.0
						Radium 226	SW9315 E903.0 Radium 228 SW9320 E904.0
						Strontium 89	D5811-95
						Gamma Isotopes **	E901.1

Comments:
 * Drew Pitzer
 Elaine Aquista
 Gary Tripp
 (circle one) EST CST MST PST
 ** Rush gamma screen - then do alpha spec for isotopic Th, U, Pb-210, Po-210
 * Perform TCLP metals - arsenic, cadmium, chromium, lead, selenium, silver, vanadium, zinc, mercury, molybdenum - use alpha spec substance
 Form 2024.XIS (1/3/01)

Relinquished By: (1)
 Signature: Sally M. Cuffin
 Printed Name: Sally M. Cuffin
 Date: 4/23/04 Time: 11:00
 Company: NHEC

Relinquished By: (2)
 Signature: _____
 Printed Name: _____
 Date: _____ Time: _____
 Company: _____

Received By: (1)
 Signature: [Signature]
 Printed Name: Amey Wolf
 Date: 4/23/04 Time: 1530
 Company: Paragon Analytics

Received By: (2)
 Signature: _____
 Printed Name: _____
 Date: _____ Time: _____
 Company: _____

CONDITION OF SAMPLE UPON RECEIPT FORM

CLIENT: New Horizons WORKORDER NO: 0404241

PROJECT MANAGER: Deb Fazio INITIALS: DF DATE: 4/23/04

1. Does this project require any special handling in addition to standard Paragon procedures? IS PRE-SCREENING REQUIRED? (radiochemistry, DOE, etc.)	Yes	<input checked="" type="radio"/> No
2. Are custody seals on shipping containers intact? How many custody seals are provided? _____	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes <input type="radio"/> No
3. Are the custody seals on sample containers intact?	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes <input type="radio"/> No
4. Is there a Chain-of-Custody (COC) or other representative documents, letters, or shipping memos?	<input checked="" type="radio"/> Yes	<input type="radio"/> No
5. Is the COC complete? Relinquished: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> Analyses Requested: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	N/A	<input checked="" type="radio"/> Yes <input type="radio"/> No
6. Is the COC in agreement with the samples received? No. of Samples: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> Sample ID's: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> Matrix: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> No. of Containers: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	N/A	<input checked="" type="radio"/> Yes <input type="radio"/> No
7. Were COC (if applicable) and sample labels legible?	<input checked="" type="radio"/> Yes	<input type="radio"/> No
8. Were airbills present and/or removable?	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes <input type="radio"/> No
9. Are all aqueous samples requiring chemical preservation preserved correctly (excluding volatile organics)? Are all aqueous non-preserved samples at the correct pH?	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes <input type="radio"/> No
10. Is there enough sample for requested analyses? If so, were samples placed in the proper containers?	<input checked="" type="radio"/> Yes	<input type="radio"/> No
11. Are all samples within holding times for the requested analyses?	<input checked="" type="radio"/> Yes	<input type="radio"/> No
12. Were all sample containers received intact? (not broken or leaking, etc.)	<input checked="" type="radio"/> Yes	<input type="radio"/> No
13. Are samples requiring no headspace (volatiles, reactive cyanide/sulfide, radon), headspace free? Size of bubble: ___ < green pea; ___ > green pea (List sample IDs and affected containers on Page 2)	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes <input type="radio"/> No
14. Were samples checked for and free from the presence of residual chlorine?	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes <input type="radio"/> No
15. Were the sample(s) shipped on ice?	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes <input type="radio"/> No
16. Were cooler temperatures measured at 0.1 - 6 °C? IR Gun Used*: 1 2	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes <input type="radio"/> No
17. Were all samples cooled that should have been cooled?	<input checked="" type="radio"/> N/A	<input type="radio"/> Yes <input type="radio"/> No

Cooler #'s 1
 Temperature Ambient °C
 Project Manager Signature / Date: Debba Fazio 4/26/04

A NO RESPONSE TO ANY QUESTION EXCEPT # 1 REQUIRES THE COMPLETION OF PAGE 2 OF THIS FORM

* IR Gun #1 (original): Raytek, SN SC-PM3/T29403
 IR Gun #2 (newer): Oakton, SN 2SCIR1201

PARAGON ANALYTICS
Radiochemistry Data Package

Section 9

**ADDITIONAL
SUPPORTING
DOCUMENTATION**

9

000064

Calibration Data Summary

Laboratory Name: Paragon Analytics
PAI Work Order: 0404241

Prep SOP: PAI 711
Analytical SOP: PAI 714

Reported on: Friday, May 21, 2004
9:39:58 AM

Lab Sample ID Spectrum Analysis Date	QC Type	Batch ID Analysis Run	Test Name	Detector Id	Eff Spectrum Bkg Spectrum Egy Spectrum	Eff Date Bkg Date Egy Date	RESULTS %Efficiency Bkg CPM Energy keV	FLAGS Efficiency Background Energy	LCL %Efficiency Bkg CPM Energy keV	LWL %Efficiency Bkg CPM Energy keV	UWL %Efficiency Bkg CPM Energy keV	UCL %Efficiency Bkg CPM Energy keV
0404241-3 O42413 5/19/2004	SMP	PL040513-1 PL040513-1A	Po210	2	C4051802 B4051802 C4051802	5/18/2004 5/19/2004 5/18/2004	31.36 0.0150 5541.5	Pass Pass Pass	30.08 0.0000 5500.0	30.62 0.0100 5510.0	32.81 0.1000 5590.0	33.24 0.1500 5600.0
0404241-5 O42415 5/19/2004	SMP	PL040513-1 PL040513-1A	Po210	3	C4051803 B4051803 C4051803	5/18/2004 5/19/2004 5/18/2004	32.05 0.0120 5547.9	Pass Pass Pass	30.42 0.0000 5492.4	30.96 0.0100 5502.4	33.08 0.1000 5582.4	33.62 0.1500 5592.4
0404241-7 O42417 5/19/2004	SMP	PL040513-1 PL040513-1A	Po210	4	C4051804 B4051804 C4051804	5/18/2004 5/19/2004 5/18/2004	30.72 0.0140 5551.4	Pass Pass Pass	28.41 0.0000 5514.9	28.91 0.0100 5525.3	30.80 0.1000 5566.7	31.40 0.1500 5577.1
0404241-7 O42417D 5/19/2004	DUP	PL040513-1 PL040513-1A	Po210	5	C4051805 B4051805 C4051805	5/18/2004 5/19/2004 5/18/2004	31.99 0.0200 5542.7	Warning Pass Pass	29.40 0.0000 5490.9	29.93 0.0100 5504.9	31.97 0.1000 5580.9	32.50 0.1500 5590.9
0404241-9 O42419 5/19/2004	SMP	PL040513-1 PL040513-1A	Po210	7	C4051807 B4051807 C4051807	5/18/2004 5/19/2004 5/18/2004	31.80 0.0140 5548.8	Pass Pass Pass	30.43 0.0000 5497.0	30.97 0.0100 5507.0	32.99 0.1000 5587.0	33.63 0.1500 5597.0
0404241-11 O424111 5/19/2004	SMP	PL040513-1 PL040513-1A	Po210	8	C4051808 B4051808 C4051808	5/18/2004 5/19/2004 5/18/2004	31.59 0.0090 5533.4	Pass Warning Pass	29.87 0.0000 5507.4	30.40 0.0100 5517.4	32.48 0.1000 5597.4	33.01 0.1500 5607.4
0404241-12 O424112 5/19/2004	SMP	PL040513-1 PL040513-1A	Po210	49	C4051849 B4051849 C4051849	5/18/2004 5/19/2004 5/18/2004	28.66 0.0140 5544.8	Pass Pass Pass	27.23 0.0000 5495.0	27.71 0.0100 5504.0	29.61 0.1000 5585.0	30.10 0.1500 5595.0
PL040513-1 O40511B 5/19/2004	MB	PL040513-1 PL040513-1A	Po210	50	C4051850 B4051850 C4051850	5/18/2004 5/19/2004 5/18/2004	31.41 0.0200 5553.5	Pass Pass Pass	29.12 0.0000 5496.0	29.64 0.0100 5506.0	31.57 0.1000 5586.0	32.18 0.1500 5596.0
PL040513-1 O40511L 5/19/2004	LCS	PL040513-1 PL040513-1A	Po210	51	C4051851 B4051851 C4051851	5/18/2004 5/19/2004 5/18/2004	30.94 0.0190 5553.0	Pass Pass Pass	29.23 0.0000 5493.3	29.75 0.0100 5503.3	31.69 0.1000 5583.3	32.31 0.1500 5593.3

Data Package ID: Po0404241-1

Abbreviations: Eff - Efficiency Bkg - Background UWL - Upper Warning Limit
Egy - Energy CPM - Counts per Minute LWL - Lower Warning Limit
UCL - Upper Control Limit CI - The Analysis Date exceeds the Calibration Date by more than 7 days.

Date Printed: Friday, May 21, 2004

Paragon Analytics

LIMS Version: 5.018A

Page 1 of 1

000065

Alpha Spectroscopy

Quality Control Data

Weekly Background, Energy, and
Efficiency Calibrations

Alpha Spec Calibration Source Re-Certification

RAINSTALPHAICLRTS04.XLS

Primary Certified Source

Source PAI ID 190 was recalibrated by Isotope Products Laboratories on 03-01-2003 and received by PAI on 03-04-2003.

Source ID: 92MIX223027; PAI ID 190 (Labeled #9)			
Total Activity:	3754 dpm		
Ref. Date:	3/1/03		
Count Date:	3/22/04		
U-234 Activity:	79.06%	=	2967.90 dpm (decay corrected)
Am-241 Activity:	19.20%	=	719.56 dpm (decay corrected)
Combined Activity:		=	3687.46 dpm (decay corrected)

Detector 13 Efficiency Determination

Source Serial #	PAI ID	Sequential #	Count Date	Am-241 net cts	U-234 net cts	count dur (s)	Combined Known cpm	Known dpm	detector efficiency
92MIX223027	190	97-19-103-09	3/22/04	7824.65	32919.75	2100	1164.126	3687.46	31.57%

Sources 1 through 8 activity determination

Source Serial #	PAI ID	Sequential #	Count Date	Am-241 net cts	U-234 net cts	count dur (s)	detector efficiency	Am-241 dpm	U-234 dpm	combined dpm
92MIX2203026	182	97-19-103-01	3/22/04	13674.65	81078.76	2100	31.57%	1237.59	7337.81	8575.40
92MIX2203028	183	97-19-103-02	3/22/04	15497.65	153089.76	2100	31.57%	1402.57	13864.97	15257.54
92MIX2203024	184	97-19-103-03	3/22/04	72039.65	74346.76	2100	31.57%	6519.75	6728.55	13248.30
92MIX2203021	185	97-19-103-04	3/22/04	22309.65	63564.76	2100	31.57%	2019.07	5752.75	7771.83
92MIX2203025	186	97-19-103-05	3/22/04	102504.65	128055.76	2100	31.57%	9276.90	11408.33	20685.23
92MIX2203022	187	97-19-103-06	3/22/04	77856.69	83352.76	2100	31.57%	7028.11	7543.61	14571.72
92MIX2203023	188	97-19-103-07	3/22/04	48378.65	70580.76	2100	31.57%	4197.37	6387.72	10585.09
92MIX2203029	189	97-19-103-08	3/22/04	34881.65	219992.76	2100	31.57%	3156.87	19909.84	23066.71

Detector 13 Efficiency Verification

Source Serial #	PAI ID	Sequential #	Count Date	Am-241 net cts	U-234 net cts	count dur (s)	Combined Known cpm	Known dpm	detector efficiency	% difference from 1st count
92MIX223027	190	97-19-103-09	3/22/04	7548.89	32241.76	2100	1136.813	3687.46	30.83%	2.35%

Sources 1 through 8 activity re-verification

Source Serial #	PAI ID	Sequential #	Combined Observed dpm	Combined Certified dpm*	Percent Difference %	Within 5% of Certified value Yes/No
92MIX2203026	182	97-19-103-01	8575.40	8730.07	1.77%	Yes
92MIX2203028	183	97-19-103-02	15257.54	15767.93	3.24%	Yes
92MIX2203024	184	97-19-103-03	13248.30	13517.34	1.99%	Yes
92MIX2203021	185	97-19-103-04	7771.83	8130.72	4.41%	Yes
92MIX2203025	186	97-19-103-05	20685.23	20951.92	1.27%	Yes
92MIX2203022	187	97-19-103-06	14571.72	15242.25	4.40%	Yes
92MIX2203023	188	97-19-103-07	10585.09	10755.77	1.59%	Yes
92MIX2203029	189	97-19-103-08	23066.71	23263.22	0.84%	Yes

*Sources 185, 186, 187, & 188 decay corrected to 04/01/03.

*Sources 182, 183, 184, & 189 decay corrected to 05/01/03.

OK - RG
EXP 3/22/05

000067



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

α 1

PAI 1807
recalibrated 4-15-03

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203026

Contained Radioactivity:

U-234:	3.354 nCi (124.1 Bq)	Am-241:	0.6793 nCi (21.43 Bq)
U-235:	0.06566 nCi (2.429 Bq)	Total Activity:	3.999 nCi (148.0 Bq)

Physical description:

A. Capsule type:	Disk (22 mm OD X 0.79 mm THK)
B. Nature of active deposit:	Electrodeposited and diffusion bonded oxides
C. Active Diameter:	19 mm
D. Backing:	Stainless steel
E. Cover:	None

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Mar 1998.

Uncertainty of Measurement:

A. Type A (random) uncertainty:	$\pm 0.7\%$
B. Type B (systematic) uncertainty:	$\pm 3.0\%$
C. Uncertainty in aliquot weighing:	$\pm 0.0\%$
D. Total uncertainty at the 99% confidence level:	$\pm 3.1\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 4483 α /min in 2 π on 11 Apr 03.

Donald James...
Quality Control

15-May-03
Date Signed

IPL Ref. No.: 987-7

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burnham, California 91504

030068

**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

22

*PAI 183
Recalibrated 4-15-03*

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203028

Contained Radionctivity:
U-234: 6.467 nCi (239.3 Bq)
U-235: 0.1135 nCi (4.200 Bq)
Am-241: 0.6366 nCi (23.55 Bq)
Total Activity: 7.217 nCi (267.1 Bq)

Physical description:

- | | |
|------------------------------|--|
| A. Capsule type: | Disk (22 mm OD X 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and-diffusion bonded oxides |
| C. Active Diameter: | 19 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

- | | |
|---|-------------|
| A. Type A (random) uncertainty: | $\pm 0.7\%$ |
| B. Type B (systematic) uncertainty: | $\pm 3.0\%$ |
| C. Uncertainty in aliquot weighing: | $\pm 0.0\%$ |
| D. Total uncertainty at the 99% confidence level: | $\pm 3.1\%$ |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 8091 α /min in 2π on 11 Apr 03.

David James Van Dalsen
Quality Control

15-May-03
Date Signed

IPL Ref. No.: 987-7

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts, Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street, Burbank, California 91504

000065



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

α3

*PAT I.D 184
recalibrated 4-15-03*

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A:	U-234	Customer:	PARAGON ANALYTICS, INC.
Radionuclide B:	U-235	P.O. No.:	EW040203/R2193
Radionuclide C:	Am-241	Catalog No.:	MISC-STD
Half Life (U-234):	(2.454 ± 0.006)E+05 years	Reference Date:	1-May-03 12:00 PST
Half Life (U-235):	(7.037 ± 0.011)E+08 years	Source No.:	92MIX2203024
Half Life (Am-241):	432.17 ± 0.66 years		

Contained Radioactivity:		Am-241:	2.866 nCi (108.0 Bq)
U-234:	3.227 nCi (119.4 Bq)	Total Activity:	6.145 nCi (227.3 Bq)
U-235:	0.05205 nCi (1.926 Bq)		

Physical description:

- | | |
|------------------------------|--|
| A. Capsule type: | Disk (22 mm OD. X 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and diffusion bonded oxides |
| C. Active Diameter: | 18 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radioimpurities: Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

- | | |
|---|--------|
| A. Type A (random) uncertainty: | ± 0.6% |
| B. Type B (systematic) uncertainty: | ± 3.0% |
| C. Uncertainty in aliquot weighing: | ± 0.0% |
| D. Total uncertainty at the 99% confidence level: | ± 3.1% |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 6889 α/min in 2π on 11 Apr 03.

[Signature]
Quality Control

15-407-013
Date Signed

IPL Ref. No.: 987-7

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504

000076



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

24

PAID TO C O I S
Rec'd from recalibrator
3-25-03

**CERTIFICATE OF CALIBRATION
ALPHA STANDARD SOURCE**

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW030603/R2155
Catalog No.: MISC-STD
Reference Date: 1-Apr-03 12:00 PST
Source No.: 92MIX2203021

Contained Radioactivity:

U-234:	2.731 nCi (101.0 Bq)	Am-241:	0.9325 nCi (34.50 Bq)
U-235:	0.03416 nCi (1.264 Bq)	Total Activity:	3.698 nCi (136.8 Bq)

Physical description:

- | | |
|------------------------------|--|
| A. Capsule type: | Disk (22 mm OD X 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and diffusion bonded oxides |
| C. Active Diameter: | 19 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

- | | |
|---|-------------|
| A. Type A (random) uncertainty: | $\pm 0.8\%$ |
| B. Type B (systematic) uncertainty: | $\pm 3.1\%$ |
| C. Uncertainty in aliquot weighing: | $\pm 0.0\%$ |
| D. Total uncertainty at the 99% confidence level: | $\pm 3.2\%$ |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 4145 α /min in 2π on 18 Mar 03.

Daniel J. Dennis
Quality Control

19-Mar-03
Date Signed

IPL Ref. No.: 987-2

ISO 9001 CERTIFIED

Medical Imaging Laboratory
1927 Avenue Tibbitts, Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street, Burbank, California 91504

000071



**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

25

41 E-D 00-186
recalibration
received 196
3-28-03

**CERTIFICATE OF CALIBRATION
ALPHA STANDARD SOURCE**

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW030603/R2155
Catalog No.: MISC-STD
Reference Date: 1-Apr-03 12:00 PST
Source No.: 92MIX2203025

Contained Radioactivity:

U-234: 5.486 nCi (203.0 Bq)
U-235: 0.09221 nCi (3.412 Bq)

Am-241: 3.956 nCi (146.4 Bq)
Total Activity: 9.536 nCi (352.8 Bq)

Physical description:

- A. Capsule type:
- B. Nature of active deposit:
- C. Active Diameter:
- D. Backing:
- E. Cover:

Disk (22 mm OD X 0.79 mm THK)
Electrodeposited and diffusion bonded oxides
19 mm
Stainless steel
None

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

- A. Type A (random) uncertainty: $\pm 0.8\%$
- B. Type B (systematic) uncertainty: $\pm 3.1\%$
- C. Uncertainty in aliquot weighing: $\pm 0.0\%$
- D. Total uncertainty at the 99% confidence level: $\pm 3.2\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 10690 α /min in 2π on 18 Mar 03.

Daniel Kenneth Dan Dalson
Quality Control

19-Mar-03
Date Signed

IPL Ref. No.: 987-2

ISO 9001 CERTIFIED

Medical Imaging Laboratory
24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory
1800 North Keystone Street Burbank, California 91504

000072

**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24917 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

26

DATE, CC 187
level for recalibrate
3-28-03

CERTIFICATE OF CALIBRATION ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW030603/R2155
Catalog No.: MISC-STD
Reference Date: 1-Apr-03 12:00 PST
Source No.: 92MIX2203022

Contained Radioactivity:

U-234: 3.592 nCi (132.9 Bq)
U-235: 0.08556 nCi (3.166 Bq)

Am-241: 3.279 nCi (121.3 Bq)
Total Activity: 6.957 nCi (257.4 Bq)

Physical description:

- | | |
|------------------------------|--|
| A. Capsule type: | Disk (22 mm OD X 0.79 mm THK) |
| B. Nature of active deposit: | Electrodeposited and diffusion bonded oxides |
| C. Active Diameter: | 19 mm |
| D. Backing: | Stainless steel |
| E. Cover: | None |

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:

- | | |
|---|--------|
| A. Type A (random) uncertainty: | ± 0.8% |
| B. Type B (systematic) uncertainty: | ± 3.1% |
| C. Uncertainty in aliquot weighing: | ± 0.0% |
| D. Total uncertainty at the 99% confidence level: | ± 3.2% |

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 7799 α /min in 2 π on 18 Mar 03.

Daniel James Van Der Kolk
Quality Control

19-Mar-03
Date Signed

IPL Ref. No.: 987-2

ISO 9001 CERTIFIED

000073

**Isotope Products
Laboratories**

An Eckert & Ziegler Company

24937 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

27

PAID 158
Spec for recalibration
3-28-03

CERTIFICATE OF CALIBRATION ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW030603/R2155
Catalog No.: MISC-STD
Reference Date: 1-Apr-03 12:00 PST
Source No.: 92MIX2203023

Contained Radioactivity:
U-234: 2.895 nCi (107.1 Bq) Am-241: 1.953 nCi (72.26 Bq)
U-235: 0.02502 nCi (0.9257 Bq) Total Activity: 4.873 nCi (180.3 Bq)

Physical description:
A. Capsule type: Disk (22 mm OD X 0.79 mm THK)
B. Nature of active deposit: Electrodeposited and diffusion bonded oxides
C. Active Diameter: 19 mm
D. Backing: Stainless steel
E. Cover: None

Radioimpurities: Not determined

Method of Calibration:
This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Aug 1992.

Uncertainty of Measurement:
A. Type A (random) uncertainty: $\pm 0.8\%$
B. Type B (systematic) uncertainty: $\pm 3.1\%$
C. Uncertainty in aliquot weighing: $\pm 0.0\%$
D. Total uncertainty at the 99% confidence level: $\pm 3.2\%$

- Notes:
- See reverse side for leak test(s) performed on this source.
 - IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
 - Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1966.
 - This source has a working life of 2 years.
 - This source had a total alpha surface emission rate of 5463 α /min in 2 π on 18 Mar 03.

Daniel Thomas Van Dalen
Quality Control

19-Mar-03
Date Signed

IPL Ref. No.: 987-2

**Isotope Products
Laboratories**

An Eckert & Ziegler Company

14917 Avenue Tibbitts
Valencia, California 91355

Tel 661-309-1010
Fax 661-257-8303

28

PAI ID 189
rec'd 4-21-03
recalibrated 4-15-03

CERTIFICATE OF CALIBRATION MIXED ALPHA STANDARD SOURCE

Radionuclide A: U-234
Radionuclide B: U-235
Radionuclide C: Am-241
Half Life (U-234): $(2.454 \pm 0.006)E+05$ years
Half Life (U-235): $(7.037 \pm 0.011)E+08$ years
Half Life (Am-241): 432.17 ± 0.66 years

Customer: PARAGON ANALYTICS, INC.
P.O. No.: EW040203/R2193
Catalog No.: MISC-STD
Reference Date: 1-May-03 12:00 PST
Source No.: 92MIX2203029

Contained Radioactivity:

U-234: 9.048 nCi (334.8 Bq)
U-235: 0.1771 nCi (6.553 Bq)

Am-241: 1.433 nCi (53.02 Bq)
Total Activity: 10.66 nCi (394.4 Bq)

Physical description:

- A. Capsule type: Disk (22 mm OD X 0.79 mm THK)
- B. Nature of active deposit: Electrodeposited and diffusion bonded oxides
- C. Active Diameter: 19 mm
- D. Backing: Stainless steel
- E. Cover: None

Radioimpurities:

Not determined

Method of Calibration:

This source was assayed using a windowless internal gas flow proportional counter for total alpha activity. Individual nuclide ratios were taken from those determined in Mar 1998.

Uncertainty of Measurement:

- A. Type A (random) uncertainty: $\pm 0.5\%$
- B. Type B (systematic) uncertainty: $\pm 3.0\%$
- C. Uncertainty in aliquot weighing: $\pm 0.0\%$
- D. Total uncertainty at the 99% confidence level: $\pm 3.0\%$

Notes:

- See reverse side for leak test(s) performed on this source.
- IPL participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (As in NRC Regulatory Guide 4.15).
- Nuclear data was taken from "Table of Radioactive Isotopes", edited by Virginia Shirley, 1986.
- This source has a working life of 2 years.
- This source had a total alpha surface emission rate of 11950 α /min in 2π on 11 Apr 03.

David J. ...
Quality Control

15-May-03
Date Signed

IPL Ref. No.: 987-7

ISO 9001 CERTIFIED

Medical Imaging Laboratory
14917 Avenue Tibbitts, Valencia, California 91355

Industrial Gauging Laboratory
1800 North Covington Street, Burbank, California 91504

000075